Wake Forest Baptist Health	**Standard Operating Procedure (SOP) Template**	

ULTRAVIOLET (UV) LIGHTS

Effective Date:	**12/15/2011**	**Revised Date:**	**2/1/2012**

INTRODUCTION

A Standard Operating Procedure (SOP) describes how your lab will handle a hazardous chemical safely, including the amount and concentration you will use, how you obtain or create the working solution, and special handling procedures, engineering controls, and personal protective equipment. Chemical-specific SOPs are found on the EH&S website.

Wake Forest Baptist Medical Center has created Standard Operating Procedures (SOPs) for several chemical hazard categories and some commonly-used chemicals. The SOP provides only standard information and requires customization for each lab.

The Occupational Safety and Health Administration (OSHA) requires a written SOP for any work with hazardous chemicals in laboratories. There are additional requirements for SOPs for particularly hazardous substances, or PHSs. These SOPs are an important part of the Chemical Hygiene Plan.

GENERAL LAB RULES

1. No eating, drinking, smoking, handling contact lenses, or applying cosmetics in the laboratory.
2. Persons shall wear buttoned lab coat, long pants, safety glasses or goggles and appropriate gloves when working with hazardous chemicals.
3. Mouth pipetting is prohibited; mechanical pipetting devices are to be used at all times.
4. All procedures are performed carefully to minimize the creation of splashes or aerosols.
5. Wash hands
    - after handling chemicals materials,
    - after removing gloves, and
    - before leaving the laboratory.
6. Plastic ware should be substituted for glassware whenever possible

Add Lab Specific Rules Here

Wake Forest Baptist Health	Standard Operating Procedure (SOP) Template	

ULTRAVIOLET (UV) LIGHTS

Effective Date:	12/15/2011	Revised Date:	2/1/2012

PURPOSE

Add Lab Specific Purpose Here

PHYSICAL HAZARDS

HEALTH HAZARDS

- Biological effects from UV radiation vary with wavelength, photon energy and duration of exposure. In general, adverse effects are limited to the skin and eyes.
- Chronic exposure:
  - May accelerate the skin aging process and increase the risk of developing skin cancer.
  - May cause cataracts in humans.
  - Wavelengths in the UVA, UVB and UVC are probable human carcinogens.
- In developing gels:
  - UV protective glasses should be worn when developing gels in a UV box.

PERSONAL PROTECTIVE EQUIPMENT

EYE PROTECTION

- Safety glasses, goggles or face shields shall be worn during operations in which ULTRAVIOLET (UV) LIGHTS might contact the eyes (e.g., through vapors or splashes of solution).
- Ordinary (street) prescription glasses do not provide adequate protection. Adequate safety glasses must meet the requirements of the Practice for Occupational Education Eye and Face Protection (ANSI Z87.1-1989) and must be equipped with side shields.
- Wearing contact lenses under some circumstances provides workers with a greater choice of eye and face protection (such as goggles or full-facepiece respirators without prescription inserts) as well as better visual acuity. However, the risk is unknown for contact lens wearers compared with nonwearers working with chemicals listed in the

NIOSH Pocket Guide to Chemical Hazards [NIOSH 2004]. OSHA recommends against contact lens use when working with acrylonitrile, methylene chloride, 1, 2 dibromo-3-chloropropane, ethylene oxide, and methylene dianiline." NIOSH Publication No. 2005-139: Current Intelligence Bulletin 59, Contact Lens Use in a Chemical Environment.

HAND PROTECTION

- Use disposable nitrile gloves when working with chemicals.
- Laboratory personnel should thoroughly wash hands with soap and water before and immediately upon removal of gloves.

LAB COATS, ETC.

- Button lab coats, closed toed shoes, long pants and long sleeved clothing shall be worn when handling ULTRAVIOLET (UV) LIGHTS. Protective clothing shall be worn to prevent any possibility of skin contact with ULTRAVIOLET (UV) LIGHTS.

SAFETY SHOWER / EYEWASH

- Where the eyes or body of any person may be exposed to ultraviolet (UV) lights, suitable facilities for quick drench or flushing of the eyes and body must be provided within the work area for immediate emergency use.
- Bottle type eyewash stations are not acceptable.

DESIGNATED AREA FOR USE AND CONTAINMENT DEVICES

SPECIAL HANDLING PROCEDURES AND STORAGE REQUIREMENTS

- The bulb should be cleaned weekly with ethanol to remove dust and dirt.
- UV lamps should be checked annually to ensure the appropriate intensity of UV light is being emitted for germicidal activity (UVC). Upon request, this can be performed by Precision Air during the annual certification process.
- Overuse of UV lights, leaving on for extended periods of time, overnight or longer, can result in the germicidal wavelength not being produced.
- Keep sash closed when UV light is on.
- Minimize storage in BSC when working in the BSC and when the UV light is on.

Wake Forest Baptist Health	Standard Operating Procedure (SOP) Template	

ULTRAVIOLET (UV) LIGHTS

Effective Date:	12/15/2011	Revised Date:	2/1/2012

- UV light should not be on when working in the BSC or when the area is occupied.

Add Lab Specific Special Handling/Storage Procedures

WASTE DISPOSAL

- UV bulbs must be disposed through EH&S.
- Store UV bulbs in an unbreakable secondary container labeled with the following **"HAZARDOUS WASTE ULTRAVIOLET (UV) LIGHTS"**.
- Full containers of ULTRAVIOLET (UV) LIGHTS waste must be disposed of according to the EH&S Hazardous Waste Program.
- A Waste Ticket must be completed and submitted prior to disposing of ULTRAVIOLET (UV) LIGHTS through EH&S.

EMERGENCY PROCEDURES

Emergency Numbers:

Fire and Medical Emergencies	716-9111 (9 + 911 for Friedburg campus)
Principal Investigator's Emergency Number	Click here to enter text.
Employee Health	716-4801
Hospital Emergency Room	716-9008
Environmental Health and Safety	716-9375

Laboratory Contact Information:

Click here to enter text.	Click here to enter text.
Click here to enter text.	Click here to enter text.
Click here to enter text.	Click here to enter text.

Wake Forest Baptist Health	Standard Operating Procedure (SOP) Template	

ULTRAVIOLET (UV) LIGHTS

Effective Date:	12/15/2011	Revised Date:	2/1/2012

FIRST AID

- The nearest safety shower station is located: Insert text here
- The nearest safety eyewash is located: Insert text here

  If chemical exposure occurs, flush exposed area for 15 to 20 minutes using emergency eyewash station and/or safety shower.

- Call 716-9111 and describe the extent of injuries, chemical spilled and the amount.
- Keep people out of the area.
- Report all accidental exposures to Employee Health Services.
- Complete an online injury/illness report if there is an over-exposure to the chemical or if there is an accident involving the chemical.

SPILL AND ACCIDENT PROCEDURES

If the chemical spilled is considered a carcinogen, reproductive toxin or highly toxic chemical, contact 716-9111 and evacuate area immediately, regardless of spill amount. If you are unsure of type of chemical, contact 716-9111. For all other spills use the chart below for spill reporting and response: Chemical Spill Procedures

SPILL QUANTITY	PROPER SPILL RESPONSE
Spill less than 300 mL	**Contact Environmental Health and Safety (716-9375) and clean up spill using spill kit.**
Spill greater than 300 mL	**Do not attempt to clean up spill. Leave the Area and immediately report to WFBMC Security (716-9111).**

CERTIFICATION OF APPLICATION
PLEASE READ
THESE ITEMS AUTOMATICALLY BECOME A PART OF YOUR SOP

1. I understand that it is my responsibility to assure that all personnel working in my laboratory with any of these hazards are fully informed about their specific dangers, proper actions for safe use, steps to take in case of accidents, and are provided with all necessary safety equipment and instructions in its use.
2. I agree to follow the provisions of the Chemical Hygiene Plan.
3. I will ensure that all of my personnel attend Basic Lab Safety Training by EH&S prior to using chemical materials.
4. I agree to permit Representatives of WFSM Environmental Health and Safety to inspect the facilities where this work is being conducted.
5. All chemical waste will be disposed of through EH&S Chemical Waste Program. I understand that **drain disposal is NOT allowed**.
6. Chemical materials will be transported in closed containers.
7. Sharps and/or breakable plastic solid waste items will be placed in approved puncture-resistant container, i.e., a sharps container.
8. As soon as possible, the Chemical Occupational Hygiene Officer will be notified in writing of any proposed changes in locations where chemical materials are stored or used.
9. Additional chemicals or changes in possession limits will be requested in writing from the Chemical Occupational Hygiene Officer.
10. Chemical materials will not be transferred to other Authorized Users without prior approval of the Chemical Occupational Hygiene Officer.
11. Chemical materials will not be shipped anywhere off campus without prior approval of the Chemical Occupational Hygiene Officer.
12. Secure chemicals (including waste) to prevent unauthorized access or removal. In addition, you must control and maintain constant surveillance of chemicals that are not in storage or are in use. This can be achieved by: 1) Locking refrigerators and/or storage cabinets, 2) Locking the laboratory when no one is present, and 3) Challenging unknown persons entering the laboratory.
13. I will notify the Chemical Occupational Hygiene Officer of my intent to leave WFUHS at least **60 days** in advance. I will be responsible for disposing of my chemical materials inventory through EH&S Chemical Waste Program.

First	Last	Signature
Click here to enter text.	Click here to enter text.	
Click here to enter text.	Click here to enter text.	
Click here to enter text.	Click here to enter text.	
Click here to enter text.	Click here to enter text.	
Click here to enter text.	Click here to enter text.	
Click here to enter text.	Click here to enter text.	
Click here to enter text.	Click here to enter text.	